

(19)

Europäisches Patentamt

European Patent Office

Office européen des brevets



(11)

EP 0 920 831 A1

(12)

EUROPEAN PATENT APPLICATION

(43) Date of publication:
09.06.1999 Bulletin 1999/23

(51) Int. Cl.⁶: A61B 1/05, A61B 5/00

(21) Application number: 99103430.7

(22) Date of filing: 24.03.1995

(84) Designated Contracting States:
DE FR GB IT

(30) Priority: 28.03.1994 US 218662

(62) Document number(s) of the earlier application(s) in
accordance with Art. 76 EPC:
95912989.1 / 0 752 825

(71) Applicant:
XILLIX TECHNOLOGIES CORPORATION
Richmond, British Columbia V6V 2V4 (CA)

(72) Inventors:
• Palcic, Branko
Vancouver, British Columbia V6N 1BA (CA)

• Macaulay, Calum
Vancouver, British Columbia V5Z 3E1 (CA)
• Lam, Stephen
Vancouver, British Columbia V6T 2E3 (CA)
• Jaggi, Bruno
Vancouver, British Columbia V6K 1M8 (CA)

(74) Representative:
Spall, Christopher John
BARKER BRETTELL
138 Hagley Road
Edgbaston Birmingham B16 9PW (GB)

Remarks:

This application was filed on 22 - 02 - 1999 as a
divisional application to the application mentioned
under INID code 62.

(54) Apparatus and method for imaging diseased tissue using integrated autofluorescence

(57) An apparatus and method for imaging diseases in tissue are presented. The apparatus employs a light source for producing excitation light to excite the tissue to generate autofluorescence light and for producing illumination light to generate reflected and back scattered light (remittance light) from the tissue. Optical sensors are used to receive the autofluorescence light and the remittance light to collect an autofluorescence light image and a remittance light image. A filter acts to integrate the autofluorescence image over a range of wavelengths in which the autofluorescence intensity for normal tissue is substantially different from the autofluorescence intensity for diseased tissue to establish an integrated autofluorescence image of the tissue. The remittance light provides a background image to normalize the autofluorescence image to account for image non-uniformity due to changes in distance, angle and illumination intensity. A monitor displays the integrated autofluorescence image and the remittance light image to produce a normalized image in which diseased tissue is distinguishable from normal tissue. The optical sensor can be installed adjacent the one end of an endoscope probe inserted into a body cavity. A method for imaging diseased tissue using an integrated fluorescence image and a normalizing remittance image is also disclosed.




FIG. 2

Description

FIELD OF THE INVENTION

[0001] This invention relates to an endoscope apparatus and method for viewing tissue within the body that uses autofluorescence of the tissue to detect and delineate diseased tissue.

BACKGROUND OF THE INVENTION

[0002] Fluorescence endoscopy involves introducing excitation light into the body and collecting the emitted fluorescence light by means of a probe that is inserted into the body to the location of interest. The probe comprises a lens fitted onto a coherent bundle of glass fibres which brings the fluorescent image out of the body cavity. Alternatively, the probe can be a rigid endoscope without fibre optics. The image brought out of the body is captured by very sensitive photodetectors and further processed in an attempt to delineate diseased tissue on the basis that diseased tissue has a different fluorescence intensity than normal tissue. This process loses significant amounts of light due to the image transfer through the optical fibres and a great deal of effort in developing new fluorescence endoscope systems is directed toward new ways of acquiring, enhancing and analyzing the very faint fluorescence images to extract as much information as possible.

[0003] Other work has been done in this field that attempts to exploit the different fluorescence spectra of diseased and healthy tissue. United States Patents 5,131,398 to Alfano, United States Patent 4,930,516 to Alfano, and United States Patent 4,786,813 to Svanberg et al. disclose various equipment and methods that acquire and process fluorescence images in an attempt to detect and delineate diseased tissue.

[0004] At present, many fluorescence endoscopy systems rely on drugs with strong fluorescence signals as well as very sensitive detectors to detect the emitted fluorescence signals. Applicant has developed equipment that employs tissue autofluorescence alone, without drugs, to detect diseased tissues such as early cancer and others. In this approach, very faint fluorescence signals must be measured at two or more specific wavelengths of emitted fluorescence. To accomplish such a measurement, very sensitive detectors, such as image intensified cameras operating at the limit of their working range must be used. This arrangement works well for detecting some tumours, but care must be taken to correctly adjust the autofluorescence signals to create appropriate images with minimal noise in order to recognize the diseased site under examination. Also, image intensified cameras (or similar detectors) are very large, they employ high voltage circuitry and they cannot be made to fit the end of the endoscope. The fluorescence images must therefore be brought out of tissue cavities through the coherent optical fibers of the

endoscope before processing of the images and/or displaying them on a video monitor.

[0005] An example of such a system is described in European Patent application No. 512,965. This application assigned to Xillix Technologies Corporation, the Assignee of the present application invention, discloses a system for producing images of diseased tissue by dividing the autofluorescence light into a pair of spectral bands. A first spectral band comprises light whereby the intensity of the autofluorescence light for healthy tissue is substantially different than the intensity of light produced for diseased tissue. A second band includes light whereby the autofluorescence intensity for diseased and healthy tissue are substantially the same. Light in each spectral band is imaged and applied to a color monitor that produces a pseudo-color image of the tissue under examination.

[0006] It has been suggested that the acquired fluorescence image of endoscope systems would be of better quality if it could be collected by a sensor at the end of the endoscope probe inserted into the body. The outer diameter of a endoscope probe must be small to allow insertion into various body cavities thereby limiting the size of the sensor that can be mounted at the distal end of the apparatus. As previously stated, fluorescence images are generally extremely faint and it is not possible for these small image sensors to capture the fluorescence images. Theoretically, increasing the excitation irradiance would increase the fluorescence intensity, however, this may also result in unwanted thermal damage or photobleaching of the tissue under examination.

[0007] Prior art endoscopes have been developed that permit the image sensor to be located at the tip of the endoscope probe, however, in general, this endoscope equipment is intended for collecting reflected light and is not suitable for reliably capturing faint fluorescence images.

SUMMARY OF THE INVENTION

[0008] The present invention seeks to provide an apparatus and method that addresses the problems of the prior art by providing a new system for acquiring fluorescence images that uses integration of autofluorescence intensity over a broad band of wavelengths to acquire a well-defined autofluorescence image that is combined with a remittance light image to create a merged image in which diseased and normal tissue are readily distinguishable. The remittance light image is used to normalize the autofluorescence image to account for image non-uniformity due to changes in distance, angle and illumination intensity.

[0009] The present invention exploits applicant's discovery that in vivo tissue autofluorescence of diseased tissue is greatly reduced through a significant range of the visible light spectrum (from approximately 500nm to approximately 700nm). To exploit this natural phenome-

non, the present novel system was designed and developed to use the integrated fluorescence over this large spectral range (500 nm - 700 nm) and display the fluorescence image along with a remittance light image such that, to the observer, normal tissue appears in a different colour from diseased tissue. In the present application, remittance light refers to reflected and back scattered light. Fluorescence signal intensity is dependent on such factors as the excitation light intensity, the proximity of the excitation light source to the tissue, and the angle of incidence of the excitation light to the tissue. To normalize for these factors, prior art systems have used two (or more) different narrow wavelength bands of fluorescence, one of which is used for normalization purposes. This requires a precondition that the difference in the tissue autofluorescence intensity between the two wavelengths must not be the same as otherwise normalization of an image could not be achieved.

[0010] In the apparatus and method of the present invention, two wavelengths of light for imaging a tissue site can be used. Preferably, blue light is used as excitation light to induce tissue autofluorescence and red/near-infrared light (approx. 700 nm) is used to produce a remittance light image. Alternatively, a single wavelength of light can be used as excitation light and a remittance light image can be collected from the remitted excitation light. The integrated autofluorescence image and the remittance light image are combined and displayed in a form to allow the observer to intuitively adjust for the excitation light intensity variations.

[0011] Accordingly, in a first embodiment, the present invention provides apparatus for imaging diseases in tissue comprising:

a light source for producing excitation light to excite the tissue to generate autofluorescence light and for producing illumination light to generate remittance light from the tissue;

optical means for receiving the autofluorescence light and the remittance light to collect an autofluorescence light image and a remittance light image; means for integrating the autofluorescence image over a range of wavelengths in which the autofluorescence intensity for normal tissue is substantially different from the autofluorescence intensity for diseased tissue to establish an integrated autofluorescence image of the tissue; and

display means for displaying the integrated autofluorescence image and the remittance light image to produce a normalized image in which diseased tissue is distinguishable from normal tissue.

[0012] In a further aspect the present invention provides, a method for imaging diseases in tissue comprising:

exciting the tissue to autofluoresce with excitation

light to generate autofluorescence light;
illuminating the tissue with illumination light to generate remittance light;
detecting the autofluorescence light;
integrating the intensity of autofluorescence light over a range of wavelengths in which the autofluorescence intensity for normal tissue is substantially different from the autofluorescence intensity for diseased tissue to establish an integrated autofluorescence image of the tissue;
detecting the remittance light to establish a remittance light image; and
combining the integrated autofluorescence image and the remittance light image to produce a normalized image in which diseased tissue is distinguishable from normal tissue.

[0013] In a further embodiment, the present invention provides a fluorescence endoscope apparatus having an image sensing device that is positioned at the end of a probe comprised of light guides. The probe is introduced into the body thereby eliminating the problem of fluorescence image degradation due to light loss while travelling through the collecting light guides. The image sensing device can be positioned adjacent to the tissue surface being examined so that there is little loss of light.

[0014] To solve the problem of reliable and accurate acquisition of a faint fluorescence image the apparatus of the present invention employs an image sensing device such as a CCD (or equivalent such as CID) array having a sensitivity that can be varied. A preferred way to achieve increased sensitivity of the CCD array is to combine individual pixel sensing elements to create larger to very large sensing units. Individual sensing elements can be combined together into 2x2, 4x4, 8x8, 16x16 or even larger groups creating new larger sensing units, particularly if CCDs are used with 100% fill factor and high quantum efficiency. With special noise suppressing circuitry very sensitive detectors can be made with very low noise, i.e. very high signal-to-noise ratio (SNR), well within parameters to detect the low fluorescence emitted by tissues.

[0015] Accordingly, in the further embodiment the present invention provides an endoscope apparatus for imaging a disease site within a body comprising:

a probe having an inner end to be located within the body near the disease site and an outer end to extend outside the body;

a light source at the outer end of the probe for generating light that includes excitation light to induce tissue fluorescence, the probe conveying the excitation light from the outer end to the inner end to illuminate the disease site and cause the tissue to fluoresce;

image sensing means at the inner end of the probe to detect tissue fluorescence, the image sensing

means having a light sensitivity that can be increased to acquire low resolution images at low fluorescent light intensities and decreased to acquire high resolution images at other light intensities;

filtering means at the inner end of the probe for filtering out the excitation light; and

display means connected to the outer end of the probe for displaying the filtered image detected by the image sensing means in such a manner as to delineate diseased and normal tissue due to diseased tissue having diminished fluorescence in comparison to normal tissue.

BRIEF DESCRIPTION OF THE DRAWINGS

[0016] Aspects of the present invention are illustrated, merely by way of example, in the accompanying drawings in which:

Figure 1a is a graph showing a typical autofluorescence curve of light intensity over a range of wavelengths for in vivo tissue exposed to excitation light; Figures 1b to 1d are graphs showing actual autofluorescence curves for tissue samples;

Figure 2 is a schematic view of a first embodiment of the present invention that employs integrated autofluorescence in which images are acquired simultaneously;

Figure 3 is a schematic view of a second embodiment of the present invention in which images are acquired sequentially;

Figure 4 is a schematic view of a third embodiment of the present invention that uses blue light to produce both the autofluorescence and remittance light images; and

Figure 5 is a schematic view of a fourth embodiment of the apparatus of the present invention that employs an CCD array positioned at the tip of an endoscopic probe.

DESCRIPTION OF THE PREFERRED EMBODIMENTS

[0017] Figure 1a shows a typical graph of the autofluorescence response spectrum of tissue in vivo when exposed to blue excitation light (400nm - 450nm). Applicant has discovered that autofluorescence of diseased tissue in vivo is greatly reduced as compared to normal tissue through the entire range of visible light (from approximately 500nm to approximately 700nm). Figures 1b-1d are examples of actual autofluorescence curves for tissue illuminated by blue excitation light. In Figure 1b, the autofluorescence spectrum for bronchial tissue is shown. Note that the autofluorescence intensity for normal tissue and tumorous tissue both peak in the green range of the spectrum (approximately 530nm), however, the peak for tumorous tissue is much lower

than the peak for normal tissue. In the region of the spectrum greater than 700nm, the spectrums for normal and tumor tissue are substantially identical. Figure 1c shows the autofluorescence spectrum for larynx tissue that is induced to fluoresce using blue laser light. Again, there is a pronounced peak in the green region of the spectrum for normal tissue and a greatly reduced spectrum for tumor tissue having a low peak in the green region. In the region of the spectrum greater than 700 nm, the spectrum is substantially the same for both tissues. In Figure 1d, the autofluorescence spectrum for nasopharynx tissue is shown having a green region peak for normal tissue and a reduced overall spectrum for diseased tissue. While there are some differences in the autofluorescence spectrum for specific types of tissue, in general, the diseased tissue autofluorescence is significantly reduced in comparison to the normal tissue autofluorescence over a broad spectral band.

[0018] Figure 2 is a schematic view of apparatus according to a first embodiment that has been developed to exploit the autofluorescence response of the tissue to differentiate between diseased and normal tissue. The apparatus of Figure 2 employs simultaneous illumination of the tissue with both excitation and illumination light. The apparatus includes a light source 30 for producing excitation light of a first pre-selected wavelength, preferably blue light (400-450nm), to excite the tissue to autofluoresce. The excited tissue produces a predominantly green autofluorescence light as indicated in Figures 1a-1d. Light source 30 also produces non-excitation light of a different wavelength to illuminate the tissue. The non-excitation light is preferably far red to infrared light (λ_R) including wavelengths greater than 700nm and is used to produce a remittance light image of the tissue. This non-excitation light is preferred as it falls outside the relatively broad spectral range of generally green emitted autofluorescence light ($\Delta\lambda_f$) in which autofluorescence intensity for normal tissue is substantially greater than for diseased tissue so as not to interfere with detection of the relatively faint autofluorescence image. Also, far red to infrared light has a low absorption coefficient in most tissue and it has a large scattering coefficient. Additionally, the scattering and absorption coefficients of red to infrared light do not vary much from location to location in most tissues so that red and infrared light is almost uniformly back scattered and reflected from the tissue. This remittance light is therefore relatively uniform in its angular distribution from the tissue and the intensity of light will depend almost entirely on the illumination intensity. Areas that receive less light will backscatter proportionally less in the same manner that autofluorescence emits light proportional to the amount of excitation light. Therefore, red and infrared light can be used to generate an image to correct for non-uniformity in the autofluorescence image that might arise due to changes in the proximity of the excitation light to the tissue, the angle of incidence of the excitation light to the tissue and the intensity of the

excitation light. It should be noted that the intensity of remittance light (λ_R) is much greater than the intensity of the autofluorescence light and a vertical break in the range over which remittance light is collected is provided in Figure 1a to account for this intensity difference.

[0019] The light is delivered to the tissue under examination by a conventional endoscope probe 32 comprising a bundle of optical fibres. Probe 32 also delivers autofluorescence light and remittance light from the tissue to optical means in the form of CCD cameras 34 and 36. CCD camera 36 has associated means for integrating the intensity of the autofluorescence light over a range of wavelengths in which the autofluorescence intensity for normal tissue is substantially different from the autofluorescence intensity for diseased tissue to establish an integrated autofluorescence intensity image of the tissue. The means for integrating the intensity of autofluorescence light preferably comprises a broad band filter 49 to admit autofluorescence light over a broad range of wavelengths. In the case of blue excitation light, the resulting autofluorescence tends to be mainly green light and the intensity is integrated over the range of wavelengths from 500-650nm designated ($\Delta\lambda_4$) in Figure 1. In this manner, the need for an image intensifier is obviated as the integrated autofluorescence signal is nearly an order of magnitude stronger compared to the natural autofluorescence intensity in a narrow wave band. In addition, the detected change (DC) between normal and abnormal tissue is approximately 8-10 times, compared to the 1.5 - 2 times of prior art systems.

Referring to Figure 1, the DC is defined as:

$$DC_{\text{application}} = \frac{N}{D}$$

$$DC_{\text{priorart}} = \frac{N}{\frac{n}{d}}$$

where

- N = the autofluorescence intensity for normal tissue at a wavelength exhibiting the maximum difference between normal and diseased tissue
- D = the autofluorescence intensity for diseased tissue at a wavelength exhibiting the maximum difference between normal and diseased tissue
- n = the autofluorescence intensity for normal tissue at a wavelength exhibiting the minimum difference between normal and diseased tissue
- d = the autofluorescence intensity for diseased tissue at a wavelength exhibiting the minimum difference between normal and diseased tissue

[0020] Since n/d is usually greater than 1 (in the range of 2 to 5) in prior art systems, the detected change for the apparatus and method of the present invention is generally 2 to 5 times greater than the detected change of the prior art.

[0021] The images captured by CCD cameras 34 and 36 are processed and displayed on display means in the form of a monitor 40. The images are preferably processed in an imaging board in a computer. The integrated autofluorescence image captured by camera 36 and the remittance light image captured by camera 34 are combined to produce a normalized image in which diseased tissue is distinguishable from normal tissue. Camera 36 provides a substantially green integrated autofluorescence image that is fed to the green channel of the display and camera 34 provides a substantially red remittance light image to normalize the autofluorescence image that is fed to the red channel of the display to produce a normalized pseudo-colour image. If diseased tissue is present, the intensity of the green autofluorescence image will be reduced over the diseased area while the remittance light image will be substantially unaffected. If tissue is normal, the intensity of the green autofluorescence image will be increased relative to the diseased tissue and the remittance light image will continue to be substantially unaffected. Combining and display of the two images on monitor 40 is arranged such that the relatively constant red remittance light image is used to establish a "underlying" image that accurately displays contours in the tissue and image variations that arise due to the proximity and intensity of the light source to the tissue. The mix of the images is such that normal tissue will appear green and diseased tissue will appear red due to the reduced levels of green autofluorescence in the vicinity of the diseased tissue allowing the red remittance light to predominate.

[0022] The apparatus of Figure 2 is set up such that light source 30 illuminates tissue 42 with the excitation light and illumination light simultaneously and separate CCD cameras 34 and 36 are provided to receive the autofluorescence image and the remittance light image. In this arrangement, light splitting means in the form of dichroic mirrors 44 and 46 are provided to direct the remittance light and the autofluorescence light to the appropriate CCD camera. While traditional 50%, 50% or X%, 100-X% beam splitters may be used, dichroics increase the fluorescence image intensity available to be detected. In Figure 2, a first dichroic mirror 44 passes the red remittance light (R) to CCD camera 34 and diverts the green autofluorescence light (G) and reflected blue excitation light (B) to second dichroic mirror 46. Second dichroic mirror 46 diverts the green autofluorescence light (G) to CCD camera 36 and passes the blue excitation light (B).

[0023] Filter 49 in the path of the autofluorescence light is preferably adjustable to vary the range of wavelengths of autofluorescence light that are received by

CCD camera 36. In this manner, the "window" over which autofluorescence is integrated can be varied to fine tune the apparatus to detect different disease conditions in different tissues that tend to exhibit characteristic autofluorescence spectra as illustrated in Figures 1b to 1d. Filter 49 can be varied by inserting an appropriate filter into place for the particular tissue and the type of tumour under examination.

[0024] In addition, an optional filter means 50 can be included in the light path to prevent blue excitation light from being received by the CCD cameras.

[0025] While the apparatus of Figure 2 employs simultaneous illumination of the tissue with both excitation and illumination light, it is also possible to expose the tissue to excitation light and illumination light sequentially. Figure 3 schematically illustrates imaging apparatus according to a second embodiment of the present invention that illuminates the tissue sequentially over a timing cycle. Light source 30 illuminates tissue 42 through endoscope 32 with excitation light for a first portion of the timing cycle. Autofluorescence light is delivered via endoscope 32 and filter 60 to a single CCD camera 62 in order to capture an integrated autofluorescence image. Filter 60 is positionable before camera 62 to filter out the excitation light and pass broadband autofluorescence light. The captured autofluorescence image is transferred to image processing means and then to image storage means located in a computer 63. Light source 30 then illuminates the tissue with illumination light to generate a remittance light image. If necessary, filter 65 is moved into position to filter the remittance light. Filters 60 and 65 are mounted on a turntable arrangement 67 that is controlled by computer 63. The remittance light image is transferred to the image processing and storage means. The two images collected in the timing cycle are then combined and displayed on monitor 40 in the same manner as previously described using colour to delineate normal and diseased tissue. The cycle is then repeated under the control of computer 63.

[0026] The foregoing embodiments of the present invention employ blue excitation light and red illumination light to produce an image that distinguishes normal and diseased tissue. In certain conditions, such as when examining inflamed tissue, the foregoing systems may cause false positive readings. In order to reduce false positive readings for cancer, a third embodiment of the present invention has been developed that employs blue light as both the excitation light and the illumination light. Figure 3 is a schematic drawing of the apparatus of the third embodiment that is designed to reduce the rate of false positive readings without sacrificing true diseased tissue (true positive) detection rates.

[0027] False positive readings arise in inflamed tissue as the increased blood concentration tends to absorb blue excitation light and green autofluorescence light leading to a reduced autofluorescence signal. For prior art autofluorescence systems, the blood absorption

causes a large drop in the green autofluorescence image and a slightly less large drop in the red autofluorescence image causing the displayed image to show inflamed tissue differently from normal tissue leading to mistaken identification as diseased tissue. For the blue excitation light, red illumination light imaging of the present invention, the same problems occur. The integrated fluorescence drops and the red remittance image does not change significantly in areas of inflamed tissue, consequently, the displayed image does not appear to show normal tissue and the inflamed area is mistaken for diseased tissue.

[0028] In the apparatus of Figure 3, light source 30 provides blue light for both excitation and illumination light in a manner identical to the embodiment of Figure 2. Green autofluorescence light (G) and blue remittance light (B) are collected from the tissue site. The combined light is split at dichroic mirror 44. A blue remittance light image is captured by CCD camera 34 and a green autofluorescence image is reflected through integrating filter 49 for capture by CCD camera 49. In this arrangement, the intensity of the detected blue remittance light image tends to drop off almost as much as the green integrated fluorescence image, consequently, the displayed image of inflamed tissue does not differ significantly from normal tissue.

[0029] An additional modification of the apparatus of the third embodiment can be used to make the presence of increased blood at a tissue site appear as a recognizable colour on the displayed image that would not be confused with normal or diseased tissue. In the modified apparatus, blue excitation light and red illumination light are used and a blue remittance light image, a red remittance light image and a green integrated autofluorescence image are collected. The red and blue remittance images are captured by camera 34. Camera 34 sends the red remittance image via its red channel and the blue remittance image via its blue channel to display 40. The green integrated autofluorescence image is captured by camera 36 and sent to the green channel of the display. On the display, normal tissue appears white, inflamed tissue has a reddish hue and abnormal tissue has a purplish hue.

[0030] The technique of the present invention also allows development of an imaging system that is sufficiently small and compact to be positioned at the end of an endoscope probe. Referring to Figure 5, there is shown an endoscope system 2 for imaging a disease site within a body according to a fourth embodiment of the present invention. The apparatus includes a probe 4 having an inner end 6 that is insertable into a body to be located near the disease site. The outer end of probe 4 remains outside the body for connection to a light source 8 for generating light and display means in the form of a colour monitor 10 for displaying the image collected by probe 4.

[0031] Light source 8 generates light at the outer end of the probe that includes excitation light to induce tis-

sue fluorescence. The probe is comprised of a bundle of optical fibres 10 that convey the light of source 8 from the outer end to inner end 6 to illuminate the disease site. At inner end 6, there is provided image sensing means, preferably in the form of CCD sensor 12 to detect and collect an image of the diseased site using remittance light and emitted fluorescent light. CCD array 12 communicates with the outer end of probe 4 via electrical connection 14 that extends through probe 4.

[0032] Preferably, inner end 6 is provided with a lens 15 to focus light on CCD sensor 12. Filtering means in the form of a filter module 16 is positioned before CCD sensor 12. Filter module 16 can be fixed in place or removable remotely. Alternatively, the filtering means can be a coating applied directly to surface of CCD sensor 12. In a still further arrangement, filter module 16 can be associated with lens 15. The filter module 16 acts to exclude excitation light and pass light of other wavelengths to the image sensing means, particularly emitted fluorescent light.

[0033] The image collected by CCD sensor 12 is transmitted via line 14 to the outer end of probe 4 for viewing on display means in the form of a colour monitor 18. The image displayed on monitor 18 delineates diseased and normal tissue for viewing by the endoscope operator.

[0034] The endoscope apparatus of the fourth embodiment is able to collect the faint fluorescence images that provide information necessary for delineating diseased and normal tissue by virtue of the fact that CCD sensor 12 is provided with a light sensitivity that can be increased to acquire low resolution images at low fluorescent light intensities and decreased to acquire high resolution images at other light intensities.

[0035] The light sensitivity of CCD sensor 12 is adjusted by binning the pixels of the sensor array. Binning the pixels decreases spatial resolution. For example, a CCD sensor comprised of 250x250, (i.e. 62,500 active pixels) if binned to form 8x8 pixels per each new sensor unit would yield an effective CCD of approximately 30x30 pixels (each pixel being 64 times larger than the original pixels) providing a very low resolution image. However, dynamic binning can solve the problem of the low spatial resolution as will be described below.

[0036] In operation, light source 8 can provide a broad spectral band of excitation light ranging in wavelength between 400-480nm (broad band blue light). Such an excitation light tends to generate a strong fluorescence response in most tissues such as the lungs, the esophagus, the colon, the bladder and the skin. The peak fluorescence is in the green region at approximately 510nm and it has been shown that the fluorescence signal is greatly diminished or virtually disappears in abnormal tissue due to early cancer or dysplasia. Alternatively, the excitation light can be in a narrow spectral band of blue or violet light.

[0037] The blue excitation light of light source 8 is con-

veyed to inner probe end 6 to illuminate an internal disease site to induce maximum tissue fluorescence particularly in the green and longer wavelengths. Scattered/reflected excitation light and fluorescence light are focused by lens 15 on CCD sensor through filter module 16. Filter module 16 is selected to filter out the blue band excitation light such that only the fluorescence green and longer wavelengths (e.g. greater than 480nm) are collected by the CCD sensor. Since abnormal tissue has greatly diminished (or a lack of) fluorescence particularly in the green region, the abnormal tissue will be detected as a darker area on the image captured by the CCD sensor. The darker abnormal regions of tissue will be readily apparent on display monitor 18. The light sensitivity of the CCD sensor can be increased by binning to acquire low resolution images at low fluorescent light intensities. While the images are of reduced resolution, they are generally sufficiently detailed to accurately determine the extent and location of the diseased tissue.

[0038] The arrangement of the fourth embodiment of the present invention reduces the complexity of the apparatus as there is generally no longer the need for an image intensifier to enhance the image collected by the CCD sensor since the sensor is located immediately adjacent the image site and the binning of the CCD sensor permits acquisition of very faint fluorescence images albeit at reduced resolution.

[0039] In a variation of the fourth embodiment, high resolution non-fluorescence images and low resolution fluorescence images of the disease site can be acquired sequentially and combined for display on monitor 18. In this variation, the endoscope apparatus is identical to the fourth embodiment except light source 8 is modified to sequentially illuminate the disease site with two or more wavelengths of light over a timing cycle. One wavelength is a fluorescence excitation light that is on for the majority of the timing cycle and the other wavelength is a non-excitation light that is on for a short period of the timing cycle. This arrangement requires means to increase the image sensing means sensitivity to acquire a low resolution fluorescence image during illumination by the fluorescence excitation light and decrease the image sensing means sensitivity to acquire a high resolution image during illumination by the non-excitation light. This arrangement also requires image processing means for storing sequential images captured by the image sensing means and combining the images for display as a pseudo-colour image on the display means.

[0040] Preferably, the endoscope apparatus of the present embodiment is operated in the following manner:

[0041] Light source 8 sequentially illuminates the disease site briefly through probe 6 with red non-excitation light and blue fluorescence excitation light over a single cycle of 30msec which is consistent with video display rates. The scattered and/or reflected red light is passed

through filter module 16 in front of the CCD sensor 12. The red light is sufficiently intense that CCD sensor 12 does not have to be binned and an image of the site at high resolution is captured by the sensor. The illumination with red light is of short duration (e.g. 5msec pulse) and the image is stored in a buffer of the image processing means comprising an imaging board 20. The blue light is of much longer duration (e.g. 25 msec) to collect sufficient fluorescence light generated by the blue light. In addition, the CCD is now binned if necessary to detect and capture the fluorescence image of the site at reduced resolution. Filter module 16 excludes the blue excitation light and permits passage of green and longer wavelengths. The captured fluorescence image is fed into another buffer of the imaging board. A pseudo image is then formed by the imaging board for display on RGB monitor 18. The pseudo image is a composite image that includes a very high resolution colour image (generated using the red light) superimposed on a green to dark green lower resolution image (generated using the blue light). In the composite image, normal tissue will be green and diseased or abnormal tissue will show up as red or pink areas. This combination of fluorescence and reflectance imaging has the additional advantage of compensating for changes in fluorescence intensity due to variation in the distance and angle between the probe and the surface of the body cavity.

[0042] Although the present invention has been described in some detail by way of example for purposes of clarity and understanding, it will be apparent that certain changes and modifications may be practised within the scope of the appended claims.

Claims

1. Endoscope apparatus for imaging tissue within a body comprising:

a probe having an inner end to be located within the body near the tissue under examination and an outer end to extend outside the body;

a light source at the outer end of the probe for generating light to cause the tissue to produce autofluorescence light having intensities that are different for healthy tissue than for diseased tissue, the probe conveying the excitation light from the outer end to the inner end to illuminate the tissue;

image sensing means at the inner end of the probe to detect the autofluorescence light and produce an image of the tissue under examination on a single channel, the image sensing means having a light sensitivity that can be increased to produce low resolution images at low autofluorescence light intensities and decreased to produce higher resolution images at other light intensities;

a filter disposed between the imaging sensing means and the tissue under for filtering out the excitation light from the light that reaches the image sensing means; and

a display connected to receive a filtered image produced by the image sensing means for displaying the filtered image in such a manner as to delineate diseased and normal tissue.

2. Endoscope apparatus as claimed in claim 1 in which the image sensing means is a CCD array.

3. Endoscope apparatus as claimed in claim 2, further comprising a controller coupled to the CCD array for increasing the light sensitivity of the CCD array by binning pixels of the CCD array.

4. Endoscope apparatus as claimed in any of claims 1-3 in which the filter comprises a coating on the CCD array.

5. Endoscope apparatus as claimed in any of claims 1-3 including a lens at the inner end of the probe to focus the autofluorescence light on the image sensing means.

4. Endoscope apparatus as claimed in any of claims 1-5 in which the light source generates excitation light in a broad spectral band ranging from 400 nm to 480 nm.

5. Endoscope apparatus as claimed in any of claims 1-6 in which the light source generates excitation light in a narrow spectral band of blue or violet light.

6. Endoscope apparatus as claimed in claim 1 in which the filter comprises a long pass blue filter.

7. Endoscope apparatus as claimed in any of claims 1-6 in which the filter excludes excitation light and passes light of other wavelengths to the image sensing means.

8. Endoscope apparatus as claimed in claim 7 in which the filtering means excludes blue excitation light and passes green and longer wavelength light.

9. Endoscope apparatus as claimed in claim 7 in which the light source sequentially illuminates the tissue under examination with two or more wavelength bands of light over a timing cycle, one wavelength band being the excitation light that causes the tissue under examination to produce autofluorescence light and the other wavelength band being non-excitation light that is reflected off the tissue under examination and passed through the filter, wherein the sensitivity of the image sensing means is increased to produce a low resolution autofluorescence image of the tissue under examination during illumination by the excitation light and decreased to produce a higher resolution reflectance image of the tissue under examination during illumination by the non-excitation light.

10. Endoscope apparatus as claimed in claim 9 in which the excitation light is on for a portion of the timing cycle that is greater than a portion of the tim-

ing cycle during which the non-excitation light is on.

11. Endoscope apparatus as claimed in claim 9 or claim 10 including image processing means for storing sequential autofluorescence and reflectance images produced by the image sensing means and combining the fluorescence and reflectance images for simultaneous display as a pseudo-colour image on the display. 5

12. Endoscope apparatus as claimed in any of claims 9-11 in which the non-excitation light is substantially red and the excitation light is substantially blue. 10

15

20

25

30

35

40

45

50

55




FIG. 1a




FIG. 1b




FIG. 1c




FIG. 1d




FIG. 2




FIG. 3




FIG. 4




FIG. 5



European Patent
Office

EUROPEAN SEARCH REPORT

Application Number
EP 99 10 3430

DOCUMENTS CONSIDERED TO BE RELEVANT			
Category	Citation of document with indication, where appropriate, of relevant passages	Relevant to claim	CLASSIFICATION OF THE APPLICATION (Int.Cl.6)
Y	US 4 878 113 A (NAKAMURA KAZUNARI) 31 October 1989 * abstract * * column 4, line 60 - column 5, line 25 * * column 15, line 5 - column 18, line 4 * * figures *	1,2,4-9, 11,12	A61B1/05 A61B5/00
Y	EP 0 512 965 A (XILLIX TECHNOLOGIES CORP) 11 November 1992 * column 3, line 5 - line 42 * * column 4, line 36 - column 6, line 40 * * figures *	1,2,4-9, 11,12	
A	US 4 928 172 A (UEHARA MASAO ET AL) 22 May 1990 * abstract; figures *	1,2	
A	US 4 631 582 A (NAGASAKI TATSUO ET AL) 23 December 1986 * the whole document *	1,2	
A	US 4 768 089 A (KATO HARUO) 30 August 1988 * abstract; figure 1 *	1,2	TECHNICAL FIELDS SEARCHED (Int.Cl.6) A61B
A	EP 0 487 143 A (PHILIPS ELECTRONICS UK LTD ;PHILIPS NV (NL)) 27 May 1992 * abstract *	1,2	
The present search report has been drawn up for all claims			
Place of search THE HAGUE		Date of completion of the search 20 April 1999	Examiner Ferrigno, A
CATEGORY OF CITED DOCUMENTS X : particularly relevant if taken alone Y : particularly relevant if combined with another document of the same category A : technological background O : non-written disclosure P : intermediate document		T : theory or principle underlying the invention E : earlier patent document, but published on, or after the filing date D : document cited in the application L : document cited for other reasons & : member of the same patent family, corresponding document	

EPO FORM 1503 03/92 (P04C01)

**ANNEX TO THE EUROPEAN SEARCH REPORT
ON EUROPEAN PATENT APPLICATION NO.**

EP 99 10 3430

This annex lists the patent family members relating to the patent documents cited in the above-mentioned European search report. The members are as contained in the European Patent Office EDP file on
The European Patent Office is in no way liable for these particulars which are merely given for the purpose of information.

20-04-1999

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
US 4878113 A	31-10-1989	JP 1043228 A	15-02-1989
		JP 1948787 C	10-07-1995
		JP 6079594 B	12-10-1994
		JP 1308528 A	13-12-1989
		JP 2710386 B	10-02-1998
		JP 1297042 A	30-11-1989
		JP 2660009 B	08-10-1997
EP 0512965 A	11-11-1992	CA 2042075 A	09-11-1992
		US 5507287 A	16-04-1996
		AT 140600 T	15-08-1996
		DE 69212382 D	29-08-1996
		JP 6054792 A	01-03-1992
		US 5769792 A	23-06-1998
US 4928172 A	22-05-1990	JP 1297041 A	30-11-1989
US 4631582 A	23-12-1986	JP 2655568 B	24-09-1997
		JP 61061588 A	29-03-1986
		DE 3530778 A	20-03-1986
US 4768089 A	30-08-1988	JP 63079632 A	09-04-1988
		DE 3732436 A	07-04-1988
EP 0487143 A	27-05-1992	GB 2250089 A	27-05-1992
		DE 69107583 D	30-03-1995
		DE 69107583 T	07-09-1995
		JP 4268422 A	24-09-1992
		US 5235402 A	10-08-1993

EPO FORM P0459

For more details about this annex : see Official Journal of the European Patent Office, No. 12/82